

NUCLEAR CAMERA WITH
OPEN AND FLEXIBLE SOFTWARE ARCHITECTURE

5 This invention relates to nuclear (gamma) camera
imaging systems and, in particular, to nuclear
cameras with software architectures which are easy to
understand and readily adapted to changing needs of
clinicians.

10 Traditionally, medical image data obtained from
medical imaging systems such as nuclear cameras,
computed tomography scanners, magnetic resonance
imaging systems and ultrasound has been defined by
proprietary image data formats developed by the
various system manufacturers. Such proprietary image
15 formats mean that the images produced by the
different systems must be viewed or displayed by
proprietary viewing/display applications provided by
each vendor. A user is effectively locked into using
the particular manufacturer's proprietary formats.
20 Images produced by one imaging system cannot be
readily viewed on another manufacturer's system.
Realizing these limitations, the NEMA trade
organization has organized and defined an image
format which all manufacturers can use, known as the
25 DICOM (Digital Imaging and COMMunication in Medicine)
standard. The various manufacturers of medical
imaging equipment can provide translators by which
their proprietary formats can be converted into a
common format in which users can exchange and view
30 images on a variety of different display devices.

Data format standards such as DICOM have not
fully solved the problems of incompatibility and ease
of use, however. First of all, a user of a
particular system must have access to conversion
35 routines by which a manufacturer's proprietary image

format can be translated into the standard format. Conversion routines are not always readily available or universally successful in their operation.

5 Secondly, the proprietary image data formats are typically inflexible with respect to changes. Changes are often desirable to enable a clinician to satisfy new user requirements, or incorporate innovations into an application, use or operation of the medical imaging system, or to enable the use of
10 new software technologies that improve the performance of the medical imaging system. Thus, proprietary image formats have the effect of limiting innovation in imaging by clinicians and thwarting improvements to clinical productivity.

15 Furthermore, while image data standards such as DICOM enable users to exchange and use images from the medical imaging systems of different manufacturers, these standards are only effective when all manufacturers agree upon the standards and
20 their requirements. The need for common agreement upon standards means that changes to standards cannot be made quickly and easily. Proposed changes need to be discussed, commented upon, and agreed to by the manufacturers before being implemented for the
25 benefit of end users. Accordingly it is desirable for a medical imaging system to have a software architecture which overcomes these obstacles and readily accommodates advances in efficiencies and techniques as they are developed by the medical
30 community.

In accordance with the principles of the present invention a nuclear camera software architecture is described which addresses these deficiencies of
35 proprietary and standard formats. The software architecture of a camera of the present invention

allows the operation and performance of the camera to be readily modified by users. Modified data can be easily exchanged because of the self-descriptive nature of the software language and the ability to provide readily available format descriptions to all users. In a preferred embodiment the software architecture embraces an open format which is publicly available. In a constructed embodiment of the present invention the software architecture of the nuclear camera uses eXtensible Markup Language (XML) which is an open format that enables changes to the data format to be made by users and other manufacturers alike. The inventive software architecture presents data that is self-describing in that relationships between various pieces of data are captured in the format definition used to store and interpret the data. In a preferred embodiment, both image data and nuclear camera control information, in particular, study protocols, are defined in the open, extensible software architecture, enabling users to exchange not only new image data formats but also new system control procedures.

In the drawings:

FIGURE 1 illustrates in block diagram form a gamma camera constructed in accordance with the principles of the present invention;

FIGURE 2 illustrates the detector of a gamma camera in block diagram form;

FIGURE 3 illustrates a gamma camera protocol setup screen for a gated SPECT study;

FIGURE 4 illustrates the flow of control and image data in a gamma camera constructed in accordance with the principles of the present invention;

FIGURE 5 illustrates the organization of patient

data in a constructed embodiment of the present invention; and

FIGURES 6-9 illustrate the directory structure of the software architecture of a preferred
5 embodiment of the present invention.

Referring first to FIGURE 1, a gamma camera constructed in accordance with the principles of the present invention is shown in block diagram form. A software architecture of the present invention may be
10 used with either a single detector gamma camera, or with a dual detector gamma camera such as those shown in U.S. Patents 5,760,402 (Hug. et al.) or 6,150,662 (Hug et al.). The dual detector cameras shown in these patents are commercially available as the Forte
15 TM and SkylightTM gamma cameras from ADAC Laboratories of Milpitas, California. These camera systems include one or more detectors 10, 12 which sense scintillation events and transfer event data over a high speed serial bus 24a, 24b from each detector to
20 an acquisition control server 14. The acquisition control server 14 bins the event data into images, which are then sent by way of a router 20 to a database host 28 connected to a department Ethernet 26. The database host 28 is the computer on which
25 the acquired images are normally saved. The database host is also the processing and viewing station where 3D reconstruction, re-slicing and presentation of the processed images to the user is performed. The router 20 serves to isolate internal camera data
30 traffic from more varied departmental data traffic. A user interface personal computer 16 is coupled to the acquisition portion of the gamma camera by the router 20 and an Ethernet network 22a, 22b, 22c. The p.c. 16 controls and monitors the image data
35 acquisition and may perform additional image

processing and display functions using an image display 18. P-scope and energy spectrum data may be displayed to the user on the display 18, enabling the user to position the patient properly in front of the camera, set energy windows correctly, and review acquired data, for instance.

5 A camera detector is shown in greater detail in FIGURE 2. As is well known in the art, a gamma camera detector is composed of a collimator 32, a scintillation crystal 34, and a lightguide 36. The photons produced by the crystal 34 and guided through the lightguide 36 are received by an array of photomultiplier tubes (PMTs) 38. A scintillation event is usually received over an area covering several PMTs, and the outputs of the tubes are sensed and used to locate the position on the detector at which the radiation event was received. The PMT output signals are amplified by pre-amplifiers 42 and digitally sampled by A/D converters 44. The samples from each PMT are accumulated for the duration of a scintillation event and, since multiple PMTs are involved in the detection of a single event, the accumulated outputs of multiple PMTs are accumulated to acquire the overall energy signal for a particular scintillation event. The detected energy data and location data from each scintillation event is modified by correction circuitry 46, which produces the detector outputs for energy (E) and event location (X,Y). The event data is then binned for image processing.

10 15 20 25 30 35 The p.c. 16 and user interface display 18 allow a user via keyboard and/or pointer control to select or create a predefined set of parameters (or protocols) for direction of a SPECT imaging session or other selected study by the gamma camera system.

FIGURE 3 illustrates a parameter interface screen and configurable parameters of a nuclear camera system for data acquisition that are selected and displayed on a screen by the user. FIGURE 3 illustrates some of the parameters that are configurable by the user for a desired study. It is appreciated that once set, the configurable parameters can be saved and referenced in a computer file for subsequent recall. The stored parameters or protocol file can then be recalled and utilized for another study, thus eliminating the need to again enter the parameters for similar or identical studies. The name of the parameter file shown in FIGURE 3 is "GATED SPECT" and is indicated at 300. It is appreciated that the acquisition system, once instructed by the user, will relay the parameters set by the user to the camera in order to initialize and begin a particular study. The initiation is done by selection of processing command 357. A user interface of this type is thus versatile while at the same time providing a high degree of automation of the execution of selected study protocols.

FIGURE 4 illustrates the flow of control and image data in a gamma camera constructed in accordance with the principles of the present invention. Located on the acquisition control server hardware 14 are a control server 52 and a data server 54. These servers act to access and process data stored on an xml data storage device 56. The data stored on the device 56 includes patient schedule data 56a, protocol data 56b, isotope data 56c, and collimator data 56d. This data is used to operate and control the gamma camera 82. In a preferred embodiment this data is stored in xml format as explained more fully below. The control server uses

the data stored on the device 56 and executes programs 52a called "scripts," also stored on the xml data storage device 56, to control the operation of the gamma camera 82. The control and data servers both operate under user control through a graphical user interface (GUI) client application 60 which runs on the user interface p.c. 16.

In operation the user defines a study (exam) by associating a protocol with a particular patient through the GUI client 60. The protocol comes from the data server 54. The GUI client then sends the study to the control server 52. The protocol for the study has a script name attached to it, enabling the control server to identify the script to execute within its interpreter. The script is executed to carry out the protocol, performing calculations as needed, and can obtain certain protocol parameters from the user such as number of azimuths, time per frame, and so forth. The script also prompts the user at various points during the study for further input as needed, such as selection of a point of interest to be tracked by the detectors as they rotate about the patient.

In a constructed embodiment the scripts are written in a scripting language called EASL. EASL was developed by the assignee company of the present invention using readily available tools for writing scripting language such as LEX and YACC. YACC is a UNIX-based tool which enables a programmer to compile the grammatical terms used in EASL to interpret the scripts. The EASL program is an interpreter for the scripts, which in turn use the xml files, and is stored in the control server 52.

The event data produced by the camera detectors Head1 and Head2 in this example is processed as

described above and transferred to a binner 70. The binner interacts with the control server 52 to format image data in an xml format. The formatted image data is then further processed, transferred or stored as described more fully below.

FIGURE 5 illustrates one form of organization of an xml image file. The image file includes information about the image such as the identification 90 of the patient, the nature of the nuclear study 92 which produced the image, the series 94 of the study, and the image information 96. In the illustration this information is all contained in an xml image file 100 referred to in this example as <ADACImage.xml>. The file includes the patient and study information just described, and the image data shown as 102. The file may contain other pieces of information 112, 114 which relate to the image and may also relate to each other. The meanings of these pieces of information and their relationships are captured in a definitional file called a Document Type Definition 110 (DTD) file. The image file 100 has a pointer to the DTD file 110 by which a processor or viewer can access the DTD file and interpret the information in the image file. The pointer can be to another file or directory running on the gamma camera as described below. The pointer can also be the URL of a location where the DTD file can be found. For example, if the image is being viewed by a browser, the browser will consult the DTD at the noted URL on-the-fly, and thus be able to interpret and understand the pieces 112, 114 of information which are stored in the image file 100.

An example of an <ADACImage.xml> file is shown in Appendix 1 below. It is seen that the second line of the file contains the pointer to the DTD file

ADACImage.dtd which defines the relationships of the data elements in the xml image file. The initial portion of the file contains information of the type shown at 90-94, including the equipment and detector used, the reconstruction processing used, patient information, isotope dose, and so forth. Toward the end of the file is a block of frame pixel information.

The initial portion of a DTD file which interprets the <ADACImage.xml> file is shown in Appendix 2. This <ADACImage.dtd> file is pointed to in the second line of the <ADACImage.xml> file and provides instructions to a reader as to the interpretation of the <ADACImage.xml> file. This particular DTD file begins with information as to the purpose, creation, and revision history of the DTD file. This is followed by a definition of the various elements of the image.xml file. The <ADACImage.xml> file is written in accordance with the rules set forth in the <ADACImage.dtd> file.

The xml files are seen to be written in easily understandable grammatical terms. With xml files a user is able to make changes to the xml file structure, adding user-desired information to the data already produced by the camera system. The user is able to make changes to algorithms executed by the scripts such as study protocols carried out by the gamma camera. The grammar-based nature of the xml files makes it easy to add newly defined xml files or to change existing ones, for instance, to create custom study protocols. Image files can be modified by a hospital user, for instance to add the hospital's unique codes and information such as accession numbers to image data. The DTD file referenced by an image would in that case define the

relation of the accession number to a particular
image and patient. Since xml is an open format the
ability to make these changes is available to all
users. In most instances these changes are evident
5 in the xml files themselves due to the self-
describing nature of the xml and DTD files, with the
format defining how the different elements in the
data relate to each other. The xml-formatted files
can be viewed on a browser or on a text editor which
10 reveals the grammatical text of the files. The xml
files can also be viewed with commercially available
xml viewers such as xmlspy. The relationships of the
elements of the file can be viewed in this way as the
hierarchy of relationships is shown automatically by
15 the viewer. A particular application can thus make
use of all of the relationships defined for the data.

A portion of another example of an xml file,
<protocol.xml>, is shown in Appendix 3. The
<protocol.xml> file has a script name attached to it
20 to enable the gamma camera to carry out a study
protocol called "MIBI." Within a protocol are
various steps, each representing a different
acquisition step, e.g., for a static image, a stress
cardiac study, or a whole body study. In the
25 illustrated <protocol.xml> file the different
characteristics of the study are called "TRAITS,"
which are interpreted and defined by the
"protocol.dtd" file referred to by the <protocol.xml>
file. Typical TRAITS include such characteristics as
30 the detector rotation, the time at each detector
orientation, the frames acquired at each step, and
the isotope dosage used. The defined TRAITS are then
used in the steps of the protocol identified in the
file as <acquisitionstep>.

35 Appendix 4 illustrates an EASL script called

5 "tbbase.asl" for conducting a total body nuclear
study. The illustrated script is written in the EASL
scripting language referred to above. The script is
seen to include comment lines which make the script
easily understandable to users. The script is also
seen to control the prompts and messages displayed to
the user on the GUI as the study proceeds. In a
gamma camera of the present invention a script is
loaded into the interpreter and is used to run the
10 protocols which control the operation of the camera.

Examples of the directory hierarchy of the files
stored on the data storage device 56 are shown in
FIGURES 6-9. In these examples the directory
/export/home/atlas/ contains two types of data as
15 shown in FIGURE 6. Software shipped with the gamma
camera by the manufacturer is stored in the etc/
directory. This includes configuration files,
default protocols, default users, default isotopes,
collimators, and so forth. This is primarily data
used by the control server 52, the data server 54,
20 and the binner 70, and comprise a complete set of
files necessary to operate the gamma camera. A data/
directory is also provided for storage of data
created by the user. This includes acquired image
25 data, user defined protocols, schedules of patients,
isotopes, energy window settings, and other files
created or modified by the user.

FIGURE 7 shows further details of the
manufacturer-supplied directory etc/. The
30 collimators.xml file contains a list of all
collimators supported by the camera. Isotopes.xml
and energywindowsets.xml contain lists of isotopes
and their physical properties and default energy
window settings. Protocols.xml contains a list of
35 standard study protocols. Systems.xml contains a

list of systems used by the camera, and whether they are clients or servers. Users.xml contains a list of default users for the camera.

5 The directory data/ for user-created data is
shown in FIGURE 8. This directory has files
corresponding to those of the etc/ directory, but in
this case the files are those which have been created
by or modified by a user. For instance, a user may
take a manufacturer-supplied protocol and modify it
10 for a particular study which the user desires to
perform. The modified protocol is stored as a
protocols.xml file in the data/ directory and can be
called up and executed by the user each time the user
wants to conduct that study. The user-created data
15 takes precedence over the manufacturer-supplied data.
Any time a client asks for data from the data server
54, the server will first access the data/ directory
to see if a user-created file satisfies the request.
Only if no user-created file is found will the server
20 look for a file in the etc/ directory.

Both directories are seen to store DTD files
corresponding to the xml files, which are used to
interpret the definitions in the corresponding xml
files. The patient.dtd file, for instance, defines
25 how the patient visit and study information is
written out, that is, the format and relationship
between the various pieces of information contained
in the patients.xml file. Similarly, protocol.dtd
defines how the protocol information is written out
30 and interpreted. A DTD file is used to view an xml
file when using an xml viewer such as xmlspy, or
alternatively a text editor can be used to view the
xml file in the form shown in the appendices.

FIGURE 9 illustrates the production and storage
35 of an image.xml file by the camera. The binner 70

writes out the results of a particular study in xml
format when acquisition is completed for a given
acquisition step of a protocol. The image.xml file
is written by the binner in accordance with the rules
5 in the ADACImage.dtd file. This example shows an
image file 015N0000.xml which has been stored under a
particular patient's name, <patient_name>/, under the
data/Patients/ directory. Image.xml files are
translated with the help of an "xml2peg" translator
10 program into an image format 015N0000.img and stored
under an x2pout/ directory. This directory is
accessed by the database host 28 in response to an
external user request for the patient's images. When
a particular image has been exported to the database
15 host, the image is then stored under the
SentToDatabase/ directory and a confirmation of the
export is sent as a status update to the data server
54. Thus, images are available in both an xml format
and in an image format which is most efficient for
20 storage, transport or display by the database host.

Appendix 1 - <ADACImage.xml>

```

5      <?xml version="1.0" encoding="UTF-8"?>
      <!DOCTYPE ADACImage SYSTEM "adacimage.dtd">

      <ADACImage>
        <ADACPrivateIOD>
          <ADACAcquisitionIE numberOfXYSets="1" objectDataType="SP"
10      uniqueObjectKey="01" uniquePatientKey="" corCorrectionFlag=""
          scaleFactorOfSets="1.0" isotopeImagingMode="0"
          totalAcquisitionTime="60000" uniformityCorrection="">
            </ADACAcquisitionIE>
            <ADACEquipmentIE clientName="cameraclient">
              <ADACDetectorModule fieldOfViewInMM="597.000000"
15      leadingDetector="head1" calibrationFactor="0.145752"
              crystallThickness="0.000000" tubePatternCorection="180"
              axisOfRotationCorrection="0" angleBetweenTheDetectors="180"/>
            </ADACEquipmentIE>
            <ADACReconstructionAndProcessingIE manipulatedImage="N"
20      curveOrROISetName="" specificCurveName="" reconstructionType=""
            associatedParentFile="" associatedNormalCurveFile=""
            associatedPathAndFileName="" xDimensionOfAssociatedImage="0"
            yDimensionOfAssociatedImage="0" approximateEjectionFraction="0.0"
            associatedHistogramCurveFile=""
25      limitedReconstructionStartSlice="0">
              <FilterModule filterType="" filterOrder="0"
              filterCutoffFrequency="0.0"/>
              <ROIModule roiMode="" roiName="" roiType="" roiColor=""
30      associatedROI_ID="" roiNumberOfPoints="" centerOfBoundingBox=""
              roiBoundingBoxCoords="" xLengthOfBoundingBox=""
              yLengthOfBoundingBox="" imageFrameNumberOfROI=""
              numberOfCurvesOrROIsInSet=""/>
              <ReorientationModule reorientationType=""
              reorientationAzimuth="0" reorientationElevation="0"/>
35      <ColorModule trueColorFlag="1" associatedColorMap="0"
              customizedColorMap="" lowerLevelGrayShade="0"
              upperWindowGrayShade="0"/>
            </ADACReconstructionAndProcessingIE>
            <ADACMiscIE unused1="" unused2="" unused3="" unused4=""
40      dirEntries="" maxTimeValue="" minTimeValue="" vfrStructure=""
            rateDifference="" imageOrientation="0" sizeOfSubHeaders=""
            binningAlgoVersion="" dataStructureArray=""
            adacPatientPosition="other" programSpecifiResult=""
            verticalFOVOffsetInMM="0.000000"
45      horizontalFOVOffsetInMM="0.000000">
              </ADACMiscIE>
              <ADACCorrectionsIE>
                <CORCorrectionModule>
                  </CORCorrectionModule>
50      <ImageOffsetCorrectionModule>
                  </ImageOffsetCorrectionModule>
                </ADACCorrectionsIE>
              </ADACPrivateIOD>
            <NMImageIOD>
              <PatientIE>

```

5 <PatientModule patientsID="1122" ethnicGroup=""
patientsSex="O" patientsName="First_da26, Gary"
patientComments="" patientsBirthDate="" patientsBirthTime="">
 </PatientModule>
 </PatientIE>
 <StudyIE>
 <GeneralStudyModule studyID="015700" studyDate="20010507"
studyTime="13:25:51" accessionNumber="::200157D0J05"
10 studyInstanceID="" studyDescription="BASE-static"
referringPhysiciansName="">
 <ProcedureCodeSequence>
 <CodeSequenceMacro>
 <BasicCodedEntry/>
 <EnhancedEncodingMode/>
15 </CodeSequenceMacro>
 </ProcedureCodeSequence>
 </GeneralStudyModule>
 <PatientStudyModule occupation="" patientsAge="0"
patientsSize="0" patientsWeight="0" additionalPatientsHistory=""
20 admittingDiagnosesDescription=""/>
 </StudyIE>
 <SeriesIE>
 <GeneralSeriesModule modality="NM" seriesDate="20010507"
seriesTime="13:25:51" protocolName="" seriesNumber=""
25 operatorsName="" patientPosition="OTHER" seriesDescription=""
seriesInstanceUID="" performedProcedureStepID=""
performingPhysiciansName="" largestPixelValueInSeries="2.000000"
smallestPixelValueInSeries="0.000000"
30 performedProcedureStepStartDate=""
performedProcedureStepStartTime=""
performedProcedureStepDescription="">
 </GeneralSeriesModule>
 </SeriesIE>
 <FrameOfReferenceIE>
35 <FrameOfReferenceModule frameOfReferenceUID=""
positionReferenceIndicator=""/>
 </FrameOfReferenceIE>
 <EquipmentIE>
 <GeneralEquipmentModule stationName="Atlantis"
40 manufacturer="ADAC Laboratories" institutionName="ADAC01"
pixelPaddingValue="" spatialResolution="1.0"
deviceSerialNumber="atlas-97001" dateOfLastCalibration=""
timeOfLastCalibration="" manufacturersModelName=""
45 institutionalDepartmentName="">
 <InstitutionAddress>
 <Address zip="" city="" state="" country=""
streetName="" numberInTheStreet=""/>
 </InstitutionAddress>
 <SoftwareVersions binnerVersion="8.0b"
50 dataServerVersion="" controlServerVersion=""
graphicUserInterfaceVersion=""/>
 </GeneralEquipmentModule>
 </EquipmentIE>
 <ImageIE>
55 <GeneralImageModule imageDate="" imageTime=""
imageType="STATIC" imageComments="" instanceNumber=""
acquisitionDate="20010507" acquisitionTime="13:25:51"

```

acquisitionNumber="" patientOrientation=""
imagesInAcquisition="" qualityControlImage="N"
derivationDescription="" lossyImageCompressionRatio="">
5   </GeneralImageModule>
    <NMImagePixelModule highBit="11" bitsStored="12"
      pixelSpacing="" bitsAllocated="16" samplesPerPixel="1"
      photometricInterpretation="PALETTECOLOR"/>
    <MultiFrameModule numberOfFrames="2">
      <FrameIncrementPointer>
10    <DetectorVector> 1, 2</DetectorVector>
      </FrameIncrementPointer>
    </MultiFrameModule>
    <NMMultiFrameModule numberOfPhases="NO_VALUE_SET"
      numberOfSlices="" numberOfDetectors="2"
15    numberOfRotations="NO_VALUE_SET"
      numberOfTimeSlots="NO_VALUE_SET"
      numberOfRRIntervals="NO_VALUE_SET"
      numberOfEnergyWindows="NO_VALUE_SET">
      <FrameIncrementPointer>
20    <DetectorVector> 1, 2</DetectorVector>
      </FrameIncrementPointer>
    </NMMultiFrameModule>
    <NMImageModule imageID="" countRate="" imageType="STATIC"
      tableHeight="0.000000" tableTraverse=""
25    countsAccumulated="766.000000" processingFunction=""
      actualFrameDuration="30000" triggerSourceOrType=""
      lossyImageCompression="00">
    </NMImageModule>
    <NMIsotopeModule>
30    <EnergyWindowInformationSequence>
      <EnergyWindowRangeSequence/>
    </EnergyWindowInformationSequence>
    <RadiopharmaceuticalInformationSequence
      radiopharmaceutical="" radionuclideTotalDose="0.000000"
35    radiopharmaceuticalRoute="" radiopharmaceuticalVolume=""
      radiopharmaceuticalStopTime="" radiopharmaceuticalStartTime="">
      <CalibrationDataSequence syringeCounts=""
        energyWindowNumber="" residualSyringeCounts=""/>
      </RadiopharmaceuticalInformationSequence>
40    <RadiopharmaceuticalInformationSequence
      radiopharmaceutical="" radionuclideTotalDose="0.000000"
      radiopharmaceuticalRoute="" radiopharmaceuticalVolume=""
      radiopharmaceuticalStopTime="" radiopharmaceuticalStartTime="">
      <CalibrationDataSequence syringeCounts=""
        energyWindowNumber="" residualSyringeCounts=""/>
45    </RadiopharmaceuticalInformationSequence>
    <InterventionDrugInformationSequence
      interventionDrugDose="" interventionDrugName=""
      interventionDrugStopTime="" interventionDrugStartTime="">
50    </InterventionDrugInformationSequence>
    </NMIsotopeModule>
    <NMDetectorModule>
      <DetectorInformationSequence startAngle="0"
55    zoomCenter="" zoomFactor="0.010000" focusCenterX=""
      focusCenterY="" focalDistance="200.000000" imagePosition=""
      collimatorType="NONE" radialPosition="0.000000"
      imageOrientation="" gantryDetectorTilt=""

```



```

collimatorOrGridName="INTR" centerOfRotationOffset=""
distanceSourceToDetector="">
  <HexagonalDimension/>
  </DetectorInformationSequence>
5  </NMDetectorModule>
  <NMRReconstructionModule sliceLocation=""
sliceThickness="10" convolutionKernel="" spacingBetweenSlices=""
reconstructionDiameter=""/>
10  <ImagePixelModule rows="256" columns="256" highBit="11"
bitsStored="12" bitsAllocated="16" samplesPerPixel="1"
pixelAspectRatio="1" pixelRepresentation="0000H"
largestImagePixelValue="" redPaletteColorLUTData=""
bluePaletteColorLUTData="" smallestImagePixelValue=""
greenPaletteColorLUTData=""
15  photometricInterpretation="PALETTECOLOR"
redPaletteColorLUTDescriptor="" bluePaletteColorLUTDescriptor=""
greenPaletteColorLUTDescriptor="">
  <Frame gateNumber="0" frameNumber="1"
imageViewID="View03.Det1" timePerFrame="30000"
20  countsPerFrame="766" detectorNumber="1" maxPixelOfFrame="2"
minPixelOfFrame="0">
iVBORw0KGgoAAAANSUUhEUgAAAQAAAEAAAAAApiSv5AAAGt01EQVR4nO3
d25bbNgyFYbl5/3fuTVczY1MWDyCwAf7fRVamtSQQpHWgwM11AQAAAAAAA
AAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAAMC716bPWm+NY6wMlNftGZQHo
25  tMn5U5c7uhdkCJOsBkmT3v5/P8+w/OILOFPI49IBIBg+840xc9h35q3t+mv
jz/Ht8YS3STGRAabky3umms9rXtYWhvIgCHPZNJxBdCJbkj1oTJ2vH3MI3v
8x/zoRnS68tX8aTU+nfYFIxG22sNVVtz0Ty77SkoSZTlRqA9iW+J+9DodV0
Pq/ribsU/dqAUF312sX928miSXulPV6Ij7s1iN9k05uOmBEmbdOuSEKcikW
nqrtdQdt04Xg/cf/Rdg014EL9u/p6PX4FK/158n5smyAf4w91QtZylsMhE
30  3z425nzlFBhBIBaXRBtufn+PDrz+NVjm4fX1b0HCA7ik0mGiRismHd34S+C
KfwbvOob3m0w68JfTlmN5/sYRR8KhJZAke0q/1Ek/ZX6zAMP+7NntdVORvz
hBfyg8S9UGxceusUhsa2+TVfKq6U+Y7CKIaTODRWyAtcPx/C/L5SLb33GLp9
0Gx01/H5aB/2uLuUztT5n/D+fzJ44bIs9jXqVF71ziig0xP9757r3fT+kh9
hPsVko3uY6ww7f3Th9YfXJf+kGuLuF8yoRKgShzPhKeD8dfBp9Bnucry8Z0
35  WH4FGDD/MhVPt1Q4G7LshGq/70XpK2LGvNBQbvWOoti+w3/eomJsOn8+un3
9P2rROaVt3H/jooqbr7T1TFnes8SOHDiWlcawUy/F4dw55s0MybrnGyAvmg
l+41VmFlUpBu8/u5hiLURP3YonWrdF/f7Zm0KNrBPsw8+hm5aq7cZrEauZh
362vyMASULiMGDWGgf/3OWXe9gStNE84Qz0+aavXf5LvJCLRTqdThaFyHo
40  C3+ZlndL5Lk+koloJHC8aWf/s7BZ2W5sRCWPAroJTFnqP02pS75OCZrHXPI
0o5FSuxGvHo1uzHUat4zDAuvP2VhbvcEZ1pLHoZK3fAEa3oFOWK6z9vltX2
pjb6QxXpc08u7jwIFSq8m1WjPk4KbXkmw5wxulWGYkmn8e9N1f0XLNks+
MhmTvcahuLH0f3b8fPcV73iN5s2uF7jft5T4CEzYvIodTcb753WTqRvZgmq
/jef5ml6rvR9yNy8y+mSzi6qu80iT7TH8OzbZUOp1Vawlu4WttHXMWmHZVe
n5FVJ43AVo9dJ/tAvG17aBiLzfrbLDzrpmICZRRWnq0ctP2S5+F3+tnZvhs
45  5oXvNtL9t1GDGoPAKtiEBF9YWq+8UqS4uu6jq/92evnRUKjuVQCD7NI1fq3
jA57o3Xi8rv6xrcljVzF4DPuapu+R6cR+zYZZu7bRyzeMZr6Zg56Z9hHXin
d/Ww9Y/BEdtjJBnaprlqyeTehnPevnr/LOk0bm7bp2dcuOjkrodhMHTJIOZ
hWgt7/8De3Tmb1z1+jjCzhUNJI3Nj+VwvPgQkJh1FciGrJUylnexyArDp2p
aKzW/YXg6gMFSe2pd3zx4KZ+zTvKP7S7FSLm8KUS8as38KPfPbbvHzCVJpK
50  1f5UwQ6o2i4HpK5XT6a2ZXP9HZXmE4TPUuuvE/jWuBy3OhliRFozJ27ZIRn
z8nR/OvQKQ61T3kq151Brj12H7F3EcNZYPKulpqJTp3DyL74yaG65iE4yFI
bIAWKGRvx0MoPj8kwCJaDLLKcj75ZRzdnzpn+njBfEW7GLKTIIdLOV3ettfN
9z+WaUrVOJoEA5NORZD0AU2xs9Kbpl/OhD/dKm01F9Hn2qBXUeJe4IpTOPC
otAlgZlQaP6zkQVc65VNVtvuoRfRLyOFD3uaIp4gXWuJKzXnJS1V3tpVhP5
55  ng912Ry3yGtnr8PP38PHDZ3Y5TANlGnRdl+8qqfvPlcxtP5nGzy3u0Hgk/m
ruISxNOBSMKUfepKyelFVfUCerWlgr2rhfMeShVgVzjig38JpD8L3VSnBV/
y064NVZAU9ZhmO/r71qzcBbqNmqqoqlufsfSjNlo8G2AnjkWJ//TasMNHXH

```

ADAC19011

Appendix 2 - <ADACImage.dtd>

```

5      <!--
      file : ADACImage.dtd
      purpose: This is the DTD used to generate the XML image from the Binner
      created: 26-Oct-00
      property of: ADAC Laboratories

10      revision history:
      26-Oct-00 CM initial version
      21-Nov-00 CM i) Changed MultiframModule to MultiFrameModule
                  ii) The pixelData is moved in to the Element.
                  iii) The order of the elements are changed so that
15                      all the encoded data will appear at the very end
                      of the XML image
      04-Jan-01 CM i) Removed the Variable Frame Element
                  ii) Added the gateNumber in Frame Element
                  iii) Added the detectorNumber in Frame Element
20      25-Jan-01 CM i) Changes which will make the DTD less restrictive to
                      generate XML ouput and validation are made. These
                      include changing mandatory Elements to optional etc.
                  ii) ID's were replaced by CDATA
      15-Feb-01 CM Added the Element ADACCorrectionsIE and related elements
      16-Feb-01 CM Added the Element NameAndValue to hold data from control server
25      04-Mar-01 CM Following the fixes in the xmllib the ADACCorrectionsIE and
                      other elements were changed as optional from required.
      30-Mar-01 CM Added imageOrientation,patientRoll and adacPatientPosition
      -->
30      <!ELEMENT ADACImage (ADACPrivateIOD, NMImageIOD)>
      <!ELEMENT ADACPrivateIOD (ADACAcquisitionIE, ADACEquipmentIE,
      ADACReconstructionAndProcessingIE, ADACMiscIE, ADACCorrectionsIE)>
      <!ELEMENT NMImageIOD (PatientIE, StudyIE, SeriesIE, FrameOfReferenceIE?, EquipmentIE,
      GeneralIE?, ImageIE)>
35      <!ELEMENT ADACAcquisitionIE (StaticAcquisitionModule?, DynamicAcquisitionModule?,
      GatedAcquisitionModule?, MCDAcquisitionModule?, TotalBodyAcquisitionModule?,
      VANTAGEAcquisitionModule?)>
      <!ATTLIST ADACAcquisitionIE
40      isotopeImagingMode CDATA #IMPLIED
      uniformityCorrection CDATA #IMPLIED
      totalAcquisitionTime CDATA #IMPLIED
      objectDataType CDATA #IMPLIED
      uniquePatientKey CDATA #IMPLIED
      uniqueObjectKey CDATA #IMPLIED
      corCorrectionFlag CDATA #IMPLIED
45      numberOfXYSets CDATA #IMPLIED
      scaleFactorOfSets CDATA #IMPLIED
      >
      <!ELEMENT ADACEquipmentIE (ADACDetectorModule)>
      <!ATTLIST ADACEquipmentIE
50      clientName CDATA #IMPLIED
      >
      <!ELEMENT ADACReconstructionAndProcessingIE (FilterModule, ROIModule, ReorientationModule,
      ColorModule)>
      <!ATTLIST ADACReconstructionAndProcessingIE
55      associatedParentFile CDATA #IMPLIED
      associatedHistogramCurveFile CDATA #IMPLIED
      associatedNormalCurveFile CDATA #IMPLIED
      associatedPathAndFileName CDATA #IMPLIED
      reconstructionType CDATA #IMPLIED
60      limitedReconstructionStartSlice CDATA #IMPLIED
  
```

```

manipulatedImage CDATA #IMPLIED
curveOrROISetName CDATA #IMPLIED
specificCurveName CDATA #IMPLIED
5  xDimensionOfAssociatedImage CDATA #IMPLIED
   yDimensionOfAssociatedImage CDATA #IMPLIED
   approximateEjectionFraction CDATA #IMPLIED
   imageViewID CDATA #REQUIRED
>
10 <!ELEMENT ADACMiscIE (NameAndValue*)>
   <!ATTLIST ADACMiscIE
       sizeOfSubHeaders CDATA #IMPLIED
       dirEntries CDATA #IMPLIED
       minTimeValue CDATA #IMPLIED
       maxTimeValue CDATA #IMPLIED
15  rateDifference CDATA #IMPLIED
       dataStructureArray CDATA #IMPLIED
       vfrStructure CDATA #IMPLIED
       horizontalFOVOffsetInMM CDATA #IMPLIED
       verticalFOVOffsetInMM CDATA #IMPLIED
20  unused1 CDATA #IMPLIED
       unused2 CDATA #IMPLIED
       unused3 CDATA #IMPLIED
       unused4 CDATA #IMPLIED
       binningAlgoVersion CDATA #IMPLIED
25  programSpecificResult CDATA #IMPLIED
       imageOrientation CDATA #IMPLIED
       patientRoll CDATA #IMPLIED
       adacPatientPosition CDATA #IMPLIED
>
30 <!ELEMENT ADACCorrectionsIE (CORCorrectionModule, ImageOffsetCorrectionModule)>
   <!ELEMENT PatientIE (PatientModule)>
   <!ELEMENT StudyIE (GeneralStudyModule, PatientStudyModule?)>
   <!ELEMENT SeriesIE (GeneralSeriesModule, NMPETPatientOrientationModule?)>
   <!ELEMENT FrameOfReferenceIE (FrameOfReferenceModule)>
35 <!ELEMENT EquipmentIE (GeneralEquipmentModule)>
   <!ELEMENT GeneralIE (VOI_LUTModule?, SOPCommonModule?)>
   <!ELEMENT ImageIE (GeneralImageModule, NMImagePixelModule, MultiFrameModule,
       NMMultiFrameModule, NMImageModule, NMIsotopeModule, NMDetectorModule,
       NMTomoAcquisitionModule?, NMMultiGatedAcquisitionModule?, NMPhaseModule?,
40  NMReconstructionModule?, OverlayPlaneModule?, MultiframeOverlayModule?, CurveModule?,
       ImagePixelModule)>
   <!ELEMENT PatientModule (ReferencedPatientSequence?, OtherPatientID*, OtherPatientName*)>
   <!ATTLIST PatientModule
       patientsName CDATA #REQUIRED
45  patientsID CDATA #REQUIRED
       patientsBirthDate CDATA #REQUIRED
       patientsSex (M | F | O) #REQUIRED
       patientsBirthTime CDATA #IMPLIED
       ethnicGroup CDATA #IMPLIED
50  patientComments CDATA #IMPLIED
>
   <!ELEMENT GeneralStudyModule (PhysicianOfRecord*, NameOfPhysicianReadingStudy*,
       ReferencedStudySequence*, ProcedureCodeSequence+)>
   <!ATTLIST GeneralStudyModule
55  studyInstanceID CDATA #REQUIRED
       studyDate CDATA #REQUIRED
       studyTime CDATA #REQUIRED
       referringPhysiciansName CDATA #REQUIRED
       studyID CDATA #REQUIRED
60  accessionNumber CDATA #REQUIRED
       studyDescription CDATA #IMPLIED
>

```

5 <!ELEMENT PatientStudyModule EMPTY>
<!ATTLIST PatientStudyModule
admittingDiagnosesDescription CDATA #IMPLIED
patientsAge CDATA #IMPLIED
patientsSize CDATA #IMPLIED
patientsWeight CDATA #IMPLIED
occupation CDATA #IMPLIED
additionalPatientsHistory CDATA #IMPLIED
>
10 <!ELEMENT GeneralSeriesModule (ReferencedStudyComponentSequence?,
RequestAttributeSequence*, PerformedActionItemSequence*)>
<!ATTLIST GeneralSeriesModule
15 modality (CR | MR | US | BI | DD | ES | MA | PT | ST | XA | RTIMAGE | RTSTRUCT | RTRECORD |
DX | IO | GM | XC | CT | NM | OT | CD | DG | LS | MS | RG | TG | RF | RTDOSE | RTPLAN | HC | MG |
PX | SM) #REQUIRED
seriesInstanceUID CDATA #REQUIRED
seriesNumber CDATA #REQUIRED
laterality (R | L) #IMPLIED
20 seriesDate CDATA #IMPLIED
seriesTime CDATA #IMPLIED
performingPhysiciansName CDATA #IMPLIED
protocolName CDATA #IMPLIED
seriesDescription CDATA #IMPLIED
25 operatorsName CDATA #IMPLIED
bodyPartExamined (SKULL | CSPINE | TSPINE | LSPINE | SSPINE | COCCYX | CHEST | CLAVICLE
| BREAST | ABDOMEN | PELVIS | HIP | SHOULDER | ELBOW | KNEE | ANKLE | HAND | FOOT |
EXTREMITY | HEAD | HEART | NECK | LEG | ARM | JAW) #IMPLIED
patientPosition (HFP | HFDR | FFDR | FFP | HFS | HFDL | FFDL | FFS | OTHER) #IMPLIED
30 smallestPixelValueInSeries CDATA #IMPLIED
largestPixelValueInSeries CDATA #IMPLIED
performedProcedureStepID CDATA #IMPLIED
performedProcedureStepStartDate CDATA #IMPLIED
performedProcedureStepStartTime CDATA #IMPLIED
35 performedProcedureStepDescription CDATA #IMPLIED
>

Appendix 3 - <protocol.xml>

```
5      <?xml version="1.0" encoding="UTF-8"?>
      <!DOCTYPE protocols SYSTEM "protocol.dtd">
      <protocols>
        <!-- protocol record # 0 -->
        <protocol>
          <protoinfo>
            <protoname type="TRAIT_PROT_NAME">MIBI MIBI</protoname>
            <tval type="TRAIT_PROT_ANATOMICAL">Cardiac</tval>
          </protoinfo>
          <protosteps>
            <protostep>
              <acquisitionstep>
                <acqstepinfo>
                  <tval type="TRAIT_ACQR_SCRIPT_NAME">ect</tval>
                </acqstepinfo>
                <acqstepaction>
                  <concurrent>
                    <tval type="TRAIT_ACQR_NAME">Rest MIBI</tval>
                    <tval
type="TRAIT_ACQR_APPDATA_CLASSNAME">data.SPECTAcqStep</tval>
                    <tval type="TRAIT_ACQR_X_DIMENSION">64</tval>
                    <tval type="TRAIT_ACQR_Y_DIMENSION">64</tval>
                    <tval type="TRAIT_ACQR_ROTATION_CLOCKWISE">true</tval>
                    <tval type="TRAIT_ACQR_TOTAL_ROTATION">180</tval>
                    <tval type="TRAIT_ACQR_START_ANGLE">0</tval>
                    <tval type="TRAIT_ACQR_RELATIVE_ANGLE">90</tval>
                    <tval type="TRAIT_ACQR_DEPTH">16</tval>
                    <tval type="TRAIT_ACQR_MAX_PIXEL">65535</tval>
                    <tval type="TRAIT_ACQR_ZOOM">1</tval>
                    <tval type="TRAIT_ACQR_DETECTORS">12</tval>
                    <tval type="TRAIT_ACQR_FLOOD_TABLE">INTR.FCR</tval>
                    <tval type="TRAIT_ACQR_MILLISECONDS">10000</tval>
                    <tval type="TRAIT_ACQR_KCOUNTS">0</tval>
                    <tval type="TRAIT_ACQR_IS_GATED">>false</tval>
                    <tval type="TRAIT_ACQR_STOP_ON SATURATE">>false</tval>
                    <tval type="TRAIT_ACQR_IS_COINCIDENT">>false</tval>
                    <tval type="TRAIT_ACQR_FRAMES_PER AZIMUTH">1</tval>
                    <tval type="TRAIT_ACQR_NUM AZIMUTHS">32</tval>
                    <tval type="TRAIT_ACQR_COMBINE HEADS">>true</tval>
                    <tval type="TRAIT_ACQR_IS ECT CONTINUOUS">>false</tval>
                    <tval type="TRAIT_ACQR DOSAGE ADMIN TIME">00:00:00</tval>
                    <tval type="TRAIT_ACQR_IS ECT NONCIRCULAR">>true</tval>
                  </concurrent>
                </acqstepaction>
              </acquisitionstep>
            </protostep>
            <protostep>
              <acquisitionstep>
                <acqstepinfo>
                  <tval type="TRAIT_ACQR_SCRIPT_NAME">ect</tval>
                </acqstepinfo>
                <acqstepaction>
                  <concurrent>
                    <tval type="TRAIT_ACQR_NAME">Stress MIBI</tval>
                    <tval
type="TRAIT_ACQR_APPDATA_CLASSNAME">data.SPECTAcqStep</tval>
                    <tval type="TRAIT_ACQR_X_DIMENSION">64</tval>
                    <tval type="TRAIT_ACQR_Y_DIMENSION">64</tval>
                    <tval type="TRAIT_ACQR_ROTATION_CLOCKWISE">true</tval>
```

```

5      <tval type="TRAIT_ACQR_TOTAL_ROTATION">180</tval>
      <tval type="TRAIT_ACQR_START_ANGLE">0</tval>
      <tval type="TRAIT_ACQR_RELATIVE_ANGLE">90</tval>
      <tval type="TRAIT_ACQR_DEPTH">16</tval>
      <tval type="TRAIT_ACQR_MAX_PIXEL">65535</tval>
      <tval type="TRAIT_ACQR_ZOOM">1</tval>
      <tval type="TRAIT_ACQR_DETECTORS">12</tval>
      <tval type="TRAIT_ACQR_FLOOD_TABLE">INTR.FCR</tval>
10     <tval type="TRAIT_ACQR_MILLISECONDS">10000</tval>
      <tval type="TRAIT_ACQR_KCOUNTS">0</tval>
      <tval type="TRAIT_ACQR_IS_GATED">>false</tval>
      <tval type="TRAIT_ACQR_STOP_ON_SATURATE">>false</tval>
      <tval type="TRAIT_ACQR_IS_COINCIDENT">>false</tval>
15     <tval type="TRAIT_ACQR_FRAMES_PER_AZIMUTH">1</tval>
      <tval type="TRAIT_ACQR_NUM_AZIMUTHS">32</tval>
      <tval type="TRAIT_ACQR_COMBINE_HEADS">>true</tval>
      <tval type="TRAIT_ACQR_IS_ECT_CONTINUOUS">>false</tval>
      <tval type="TRAIT_ACQR_DOSAGE_ADMIN_TIME">00:00:00</tval>
      <tval type="TRAIT_ACQR_IS_ECT_NONCIRCULAR">>true</tval>
20     </concurrent>
      </acqstepaction>
      </acquisitionstep>
      </protostep>
      </protosteps>
25     </protocol>
      <!-- protocol record # 1 -->
      <protocol>
        <protoinfo>
          <protoname type="TRAIT_PROT_NAME">Dual Isotope</protoname>
30         <tval type="TRAIT_PROT_ANATOMICAL">Cardiac</tval>
        </protoinfo>
        <protosteps>
          <protostep>
            <acquisitionstep>
35             <acqstepinfo>
              <tval type="TRAIT_ACQR_SCRIPT_NAME">ect</tval>
            </acqstepinfo>
            <acqstepaction>
              <concurrent>
40                 <tval type="TRAIT_ACQR_NAME">ect</tval>
                 <tval

```

Appendix 4 - tbbase.asl

```
#
# file: tbbase.asl
5 # purpose: PreProgram Script for 'basic' TotalBody(tb) study.
#
# created: 3-Jan-01
# property of ADAC Laboratories
#
10 # revision history:
#     3-Jan-00      DC      initial version
#
# script dependencies:
#     - Exchanger is parked (check ColPark)
15 #     - User Override is Off
#     - Rings are locked at 180.
#
# target values (Input):
#     none
20 #
# return values (Output):
#     false - Script failed to complete
#     true - Script completed
#
25 # $Header: /usr/adac/Repository/camera/camera/control/etc/easl/scripts/tbbase.asl,v 1.18
2001/05/08 05:58:15 sheavner Exp $
#
script "tbbase"

30     # Debug message - let me know the script has started ok
     choiceNum = message( "DEBUG MESSAGE -- Starting tbbase.asl - Gantry will move
to scan setup position when you click ok")
     choices "DEBUG OK!";

35     # Discuss study with tech, subtract off one FOV
     if ( TRAIT_ACQR_SCAN_LENGTH > 0 ) {
         TRAIT_ACQR_SCAN_LENGTH = TRAIT_ACQR_SCAN_LENGTH -
TRAIT_ACQR_FOV_SIZE_Y;
     } else {
40         TRAIT_ACQR_SCAN_LENGTH = TRAIT_ACQR_SCAN_LENGTH +
TRAIT_ACQR_FOV_SIZE_Y;
     }

45     TABLE_EXTEND_POSITION_VELOCITY = 40.0;

#
# Set the starting and ending(target) position of the longitudinal
# motion and calculate the scan length, in microns for the binner.
#
50     _startTablePos = TRAIT_ACQR_TRANSLATE_POS;
     _endTablePos = TRAIT_ACQR_TRANSLATE_POS +
TRAIT_ACQR_SCAN_LENGTH;

55     _tb_lower_limit = GantryTB lower limit;
```



```

        _tb_upper_limit = GantryTB_upper_limit;
        if ( (_startTablePos > _tb_upper_limit) or ( _startTablePos < _tb_lower_limit) ) {
            choiceNum = message("EASL_STUDY_TB_START_POS",
5         _tb_lower_limit/10, _tb_upper_limit/10, ( _tb_upper_limit-
            _tb_lower_limit+TRAIT_ACQR_FOV_SIZE_Y)/10 ) choices "EASL_CONTINUE";
            return;
        }
        if ( (_endTablePos > _tb_upper_limit) or ( _endTablePos < _tb_lower_limit) ) {
10         if (TRAIT_ACQR_SCAN_LENGTH > 0) {
            _allowable_length = _tb_upper_limit-
            _startTablePos+TRAIT_ACQR_FOV_SIZE_Y;
        } else {
            _allowable_length = _tb_lower_limit-
15         _startTablePos+TRAIT_ACQR_FOV_SIZE_Y;
        }
        choiceNum = message("EASL_STUDY_TB_END_POS",
            _tb_lower_limit/10, _tb_upper_limit/10, _allowable_length/10, _startTablePos/10 ) choices
            "EASL_CONTINUE";
20         return;
        }

        #
        # Negative scan length implies that we are scanning from foot to head
        #
25         show status "*****" (script)           Starting position is " + _startTablePos + "
mm";
        show status "*****" (script)           Ending(target) position is " +
        _endTablePos + " mm";
30         show status "*****" (script)           The required scan length is " +
        TRAIT_ACQR_SCAN_LENGTH + " mm";

        # Leftover -- remove when can verify it's ok
        StudyVGAngle = 180; start StudyVGAngle;
35         # calculate the longitudinal velocity
        # ( negative indicates that we are moving from foot to head,
        #     the start position is greater than the stop position )

40         # scanSpeed is mm per second
        _scanSpeed = TRAIT_ACQR_TBODY_SPEED;

        # test the speed for a valid range

45         show status "*****" (script)           Scan Speed is " + _scanSpeed + "
mm/sec";

        # -----
        #
50         show status "%%%%%%%%%" (script)           Running TB setup script
%%%%%%%%%;
        _finishedSetup = false;
        while ( _finishedSetup = false)
        {
55         StudyTBSetup = _startTablePos;
            run StudyTBSetup;

```

```

    if (StudyTBSetup != true)
    {
        if (StudyTBSetup < 0)
            return;
5         choicenum = message("EASL_STUDY_SETUP_FAILED") choices
"EASL_RETRY", "EASL_CANCEL";
        if (choicenum = 2)
            return;
    }
10        else
        {
            _finishedSetup = true;
        }
    }
15
# -----
#

20        # Start binner and wait for the start button
        show status " %%%%%%%%%%(script)          WAITING FOR START
%%%%%%%%%%";

        choicenum = message( "DEBUG MESSAGE -- Gantry is in position, Click start to
25        begin study (after clicking ok)" )
            choices "OK";
        # status message( "Press the START button to begin");

        enable imaging;
30        wait until user requests imaging start;

        show status " %%%%%%%%%%(script)          User has pressed START
%%%%%%%%%%";
35        # -----
        #

        #
        # For Total Body acquisitions there are 3 stages
        #
40        # 1. The mask is opened one row at a time and when fully open the binning halts
        waiting for the table to start moving
        # 2. The table moves the required distance at the required speed with the mask fully
        open and binning in progress
        # 3. The table stops moving and the mask automatically starts to close.
45        #

        #
        # Setup the target position for the acquisition and the speed for the table motion during
        the acquisition
50        #
        StudyTB = _endTablePos;
        StudyTB velocity = _scanSpeed;

        # Start StudyMonitor to watch for gantry exceptions, start after setup.ppm completes
55        start StudyMonitor;

```

```

#
# send the start imaging tag into the data stream and so to the
# binner to start the imaging
#
5      start imaging;

      show status "%%%%%%%%%(script)          Imaging has started, waiting for
the mask to be fully open %%%%%%%%%%";

10     # Wait for the mask to be fully open. This is indicated by the binner state changing
from
      # Inprogress to ReadyToMove

15     #
      # check that the imaging did start
      #
      _loop = 90;
      while (imaging is not underway)
20     {
          if ( _loop < 0 )
          {
              show status "%%%%%%%%%(script) Imaging did not start
after 90 seconds";
25     choiceNum = message( "DEBUG MESSAGE -- Imaging did not
start after 90 seconds." )
              choices "Abort";
              stop StudyMonitor;
              current setting = false;
              return;
          }
          _loop = _loop - 1;
          pause 1.0;
35     }

      show status "%%%%%%%%%(script)          Imaging has started, waiting for
the mask to be fully open %%%%%%%%%%";

40     # Wait for the mask to be fully open. This is indicated by the binner
# state changing from Inprogress to ReadyToMove

      _quiet = 0;
      while (imaging is not ready for motion)
45     {
          if ( StudyMonitor is not working)
          {
              show status "%%%%%%%%%(script)          Study Monitor
is not working - waiting for mask to open %%%%%%%%%%";
50     stop imaging;
              stop StudyMonitor;
              return;
          }

          if ( imaging is complete )
55     {

```

ADAC19011

ADAC19011